

PV INSULATION EARTH TESTER KEW 6024PV

No need to short circuit strings for measuring the insulation resistance!



PV Insulation

Insulation

Correct measuring for Insulation resistance

CE

IP54

even if PV arrays are under generating power Waterproof design: Can measure under the bad

weather condition

Memory function up to 1000 data

Luminescence button and Backlight


Earth

Volts

KEW 6024PV Specifications

| sulation resistance | PV Insulation* | | Insulation | | |
|----------------------------------|---|---------------|--------------------|----------------|---------------|
| Test voltage | 500V | 1000V | 250V | 500V | 1000V |
| Measuring range(Auto range) | 20.00/200.0/2000ΜΩ | | 20.00/200.0/2000ΜΩ | | |
| Mid-scale value | _ | | 50ΜΩ | | |
| Rated current | | | 1.0 - 1.2mA | | |
| | | _ | 0.25ΜΩ | 0.5ΜΩ | 1ΜΩ |
| First effective measuring range | 1.51 - 200.0MΩ | 1.51 - 1000MΩ | 1.51 - 100.0MΩ | 1.51 - 200.0MΩ | 1.51 - 1000MΩ |
| Accuracy | ±1.5%rdg ±5dgt | | ±1.5%rdg ±5dgt | | |
| Second effective measuring range | 0.00 - 1.50ΜΩ | 0.00 - 1.50ΜΩ | 1.20 - 1.50MΩ | 1.20 - 1.50MΩ | 1.20 - 1.50MΩ |
| | 200.1 - 2000ΜΩ | 1001 - 2000MΩ | 100.1 - 2000MΩ | 200.1 - 2000ΜΩ | 1001 - 2000MΩ |
| Accuracy | ±5.0%rdg ±6dgt | | | | |
| Open circuit voltage | 0 - +20% | | | | |
| Short circuit current | Max 1.5mA | | | | |
| arth resistance | | | | | |
| Measuring range(Auto range) | 20.00/200.0/2000Ω | | | | |
| Accuracy | ±3.0%rdg±0.1Ω (20Ωrange) ±3.0%rdg±3dgt (200/2000Ωrange) | | | | |
| ltage measurement | | | | | |
| Measuring range | AC 5 - 600V (45 - 65Hz) DC ±5 - ±1000V | | | | |
| Accuracy | ±1.0%rdg±4dgt | | | | |
| eneral | | | | | |
| Applicable Standards | IEC 61010-1 CAT IV 300V, CAT Ⅲ 600V Pollution2 IEC 61010-2-030, IEC 61010-031, IEC 60529(IP54), IEC 61557-1,-2,-5,-10, IEC 61326-1,2-2(EMC), EN50581(RoHS) | | | | |
| Power source | LR6(AA) (1.5V) x 6 | | | | |
| Dimension / Weight | 84(L) x 184(W) x 133(D)mm / Approx. 900g (including batteries) | | | | |
| Accessories | 7196A(Test leads with remote control switch), 7244A(Test lead with alligator clip), 8017(Extension prod), 8072(CAT II standard prod), 8212-USB(USB adaptor with "KEW Report (Software)") 9155(Shoulder strap), 9156(Carrying case), LR6 x 6, Instruction manual | | | | |
| Optional | 7243(L-shaped probe), 7245A(Precision measurement cord set), 8016(Hook type prod) | | | | |

^{*6024}PV supports the PV systems up to 1000V.






Included Accessories



Optional Accessories





Please read the "Safety Warnings" in the instruction manual supplied with the instrument thoroughly and completely **Safety Warnings**: for correct use. Failure to follow the safety rules can cause fire, trouble, electrical shock, etc. Therefore, make sure to operate the instrument on a correct power supply and voltage rating marked on each instrument.

For inquires or orders:



2-5-20, Nakane, Meguro-ku, Tokyo, 152-0031 Japan Phone:+81-3-3723-0131

Fax:+81-3-3723-0152

